

Declaration Compatibility

Manufacturer: CIMPAX ApS
Lille Pilevang 4
DK-3550 Slangerup
Denmark

Medical Device: Type: Diathermy Pencils
REF: C31-19-001, C31-19-002,
C31-19-101, C31-19-102
Classification: IIb, rule 9

We, the manufacturer, hereby declare that the above-mentioned devices has been tested according to the standards:


IEC60601-1: 2005; 3ed
IEC60601-2-2: 2009

and is guaranteed safe to use and compatible with all HF generators available including the following HF-generator brand:

Erbe VIO 300D

Place and Date of issue:

Denmark, 8. august 2016



Michael Busch-Madsen

Vice President International Sales